

Summary of Events originating in the Hospital Transfusion Laboratory - Previous Recommendations

Year first made	Action	Recommendation
2011	Transfusion Laboratory Managers, Pathology IT managers, LIMS Providers	<p>As the specification of transfusion laboratory information management systems (LIMS) is further developed it is vital that:</p> <ul style="list-style-type: none"> - As a minimum there is a requirement for positive confirmation, by the biomedical scientist (BMS), at the point of component reservation, that special requirements have been met - Preferably, a requirement for a direct check within the LIMS, that the component meets the special requirement on record - Warnings must be clear and appear on all relevant screens in the transfusion process - Warning flags need a positive response from the user as to why they are being overridden